

S-EOSTM HYPERSPECTRAL CAMERA



Integrating our new ZephIR camera, the EOS family now covers the SWIR spectral region up to 2.5 µm. Our S-EOS widefield hyperspectral imager will change your view of spectral analysis by providing complete spectral information for each and every pixel of a full resolution image. S-EOSTM delivers a series of monochromatic images, avoiding fastidious x-y or line scanning. The system allows unprecedented analysis by providing large scale distribution of spectral features, whether it is band gap variability of a semiconductor or molecular variation in a new compound.

APPLICATION EXAMPLES:

- » Photovoltaic characterization
- » Mineral analysis
- » Forensic
- » Food and plants sorting

| PERFORMANCE | | |
|---------------------------------------|--|-------------|
| STANDARD PRODUCTS | S - EOS 1.7 | S - EOS 2.5 |
| Spectral Range | 0.9 - 1.7 μm | 1 - 2.5 μm |
| Spectral Resolution | < 5 nm | |
| Spectral Width Sampling | ≥ 0.2 nm programmable | |
| Spectral Channels | Continuously tunable | |
| Spectral Image Rate | 15-20 fps | |
| Entrance Slit Size | No slit / Full field of view measured for each wavelength | |
| Pixel Size | 30 μm | |
| Dynamic Range (digitization) | 14 bits | |
| Sensor Frame Rate | Up to 346 fps | |
| Camera Type | FPA | |
| Camera Acquisition (linear or matrix) | Matrix | |
| Lens Mount Standard | C-Mount (option for CS-Mount) | |
| Cooling | Yes | |
| Camera Interface | Camera Link™ | |
| Frame Grabber Needed | Yes | |
| Exposure Control | PHySpec™ software controlled | |
| Detector Type | HgCdTe (MCT) | |
| SOFTWARE & DATA PROCESSING | | |
| Operating System | Windows 7 (64 bits) | |
| Acquisition | PHySpec™ Software | |
| Preprocessing | Image stabilization, spatial filtering, statical tools, spectrum extraction, data normalization, spectral calibration | |
| Hyperspectral Data Format | HDF5, FITS | |
| Single Image Data Format | HDF5, FITS, PNG, TIFF, JPG | |
| Spectrum Data Format | HDF5, CSV, JPG, PNG, TIFF | |
| Option | C++ SDK plugin interface included | |
| DIMENSIONS, WEIGHT & POWER | | |
| Footprint | 305 mm x 610 mm x 270 mm | |
| Weight | 20 Kg | |
| Power Consumption | ≤ 25 W (including detector) | |
| Power Supply | 24 V | |
| PORTABILITY | | |
| Mounting | 305 mm x 610 mm optical breadboard; 1/4 imperial thread | |
| Tripod | Optional | |
| ENVIRONMENTAL CONDITION | | |
| Operation Temperature | 10°C to 40°C | |
| Storage Temparature | 0°C to 50°C | |
| ACCESSORIES | | |
| Computer | Not included | |
| Reference Panels | Reflectance standard and calibration lamp | |



GRAND-EOS™ HYPERSPECTRAL CAMERA



Macro-imaging modality



| TECHNICAL SPECIFICATIONS | | | |
|---------------------------------|--|--|--|
| Spectral Range | 400 - 1620 nm | | |
| | VNIR | SWIR | |
| Spectral Resolution | < 2.5 nm (400 - 1000 nm) | < 4 nm (900 - 1700 nm) | |
| Spatial Resolution | < 7.5 μm | < 12 μm | |
| with 10 X microscope objective) | < 7.5 μπ | | |
| Camera | Front-illuminated interline CCD camera | InGaAs camera deep-cooled at -80°C | |
| Wavelength tuning speed | 60 ms stabilization time for 2 nm step | 60 ms stabilization time for 5 nm step | |
| Wavelength Absolute Accuracy | < 0.3 nm | < 0.6 nm | |
| Visualisation Camera | Monochrome or Color XMP camera - 2/3" 5.1M Progressive Color CMOS - 2448 x 2048 pixels | | |
| | | | |
| Sample Holder | XY Manual translation stage (50 mm travel) | | |
| Preprocessing | Image stabilization, spatial filtering, statical tools, spectrum extraction, | | |
| | data normalization, spectral calibration | | |
| Hyperspectral Data Format | HDF5, FITS | | |
| Single Image Data Format | HDF5, FITS, PNG, TIFF, JPG | | |
| Spectrum Data Format | HDF5, CSV, PNG, TIFF, JPG | | |
| Acquisition | PHySpec™ control and analysis software | | |
| Operating system | Windows 7 (64 bits) | | |
| Option | C++ SDK plugin interface included | | |
| | | | |
| Macro-imaging modality | | | |
| Field of view | Optimized from 20 x 20 mm to 160 x 160 mm | | |
| Lens | 16mm focal length VIS-NIR-SWIR objective (transmittance >90% between 400 - 1700 r | | |
| | | | |
| Micro-imaging modality | | | |
| Microscope | Upright or Inverted | | |
| Objectives | 5x, 10x (other magnifications available upon request) | | |
| Illumination | Broadband and monochromatic illumination available via light guide | | |
| Excitation | 532 nm, 660 nm, 785 nm, or 808 nm lasers | | |

GRAND-EOS combines a hyperspectral microscopy system with a hyperspectral wide-field imaging platform, giving access to micro and macro modalities with both VNIR (400-1000 nm) and SWIR (900-1700 nm) spectral ranges. This imaging platform takes advantage Photon etc's patented filtering technology based on volume Bragg grating providing a non-polarized wavelength selection with high throughput and efficiency. This filtering method allows imaging of large field-of-view, scanning through a user defined wavelength range. Using a megapixel sensor, the acquisition of filtered images provides spectral information from million of points at the surface of the sample. The versatility of GRAND-EOS as well as its high spatial and spectral resolution makes it an ideal tool for both fundamental research or industrial applications.

Other wavelengths available upon request

APPLICATION EXAMPLES:

- » Photovoltaic characterization
- » Mineral analysis

- » Forensic
- » Food and plants sorting