

## V-EOS<sup>™</sup> HYPERSPECTRAL CAMERA



Our widefield hyperspectral imager will change your view of spectral analysis by providing complete spectral information for each and every pixel of a full resolution image. V-EOS™ delivers a series of monochromatic images, avoiding fastidious x-y or line scanning. The system allows unprecedented analysis by providing large scale distribution of spectral features, whether it is band gap variability of a semiconductor or molecular variation in a new compound.

## **APPLICATION EXAMPLES:**

- » Photovoltaic characterization
- » Mineral analysis
- » Forensic
- » Food and plants sorting

| PERFORMANCE                           |   |
|---------------------------------------|---|
| Spectral Range                        | 400 - 1000 nm   |
| Spectral Resolution                   | < 2.5 nm  |
| Spatial Resolution                    | 2 mm at 1.20 m with a FOV of 1.85 m x 1.35 m                  |
| Spectral Channels                     | Continuously tunable  |
| Spectral Image Rate                   | 3 fps   |
| Wavelenght Absolute Accuracy          | < 0.4 nm  |
| Spectral Width Sampling               | $\geq 0.1$ nm programmable                                    |
| Entrance Slit Size                    | No slit / Full field of view measured for each wavelength     |
| Standard Field of View (customizable) | 72° Horizontal, 50° Vertical                                  |
| Calibration                           | Automatic   |
| Pixel Size                            | 6.45 μm x 6.45 μm   |
| Dynamic Range (digitization)          | 14 bit  |
| CCD Frame Rate                        | 13.5 fps  |
| Smile Distortion                      | ≤ 0.5 nm  |
| Keystone Distortion                   | No  |
| Barrel Distortion                     | $\leq$ 2% due to widefield entrance optics                    |
| Noise level                           | 6e <sup>-</sup> @ 10 MHz                                      |
| Camera Type                           | Interline   |
| Camera Acquisition (linear or matrix) | Matrix  |
| Lens Mount Standard                   | C-Mount (option for CS-Mount)                                 |
| Camera Interface                      | Firewire  |
| Frame Grabber Needed                  | No  |
| Exposure Control                      | PHySpec <sup>™</sup> software controlled                      |
| Binning                               | 1x1; 1x2; 2x1 ; 2x2   |
| Detector Type                         | CCD   |
| Focus                                 | Motorized for Chromatism Correction                           |
| SOFTWARE & DATA PROCESSING            |   |
| Operating System                      | Windows 7 (64 bit)  |
| Acquisition                           | PHySpec <sup>™</sup> Software                                 |
| Preprocessing                         | Image stabilization, spatial filtering, statical tools,       |
|                                       | spectrum extraction, data normalization, spectral calibration |
| Hyperspectral Data Format             | FITS, HDF5  |
| Single Image Data Format              | FITS, PNG, TIFF, JPG  |
| Spectrum Data Format                  | JPG, PNG, TIFF, CSV, PDF, SGV                                 |
| Option                                | C++ SDK plugin interface included                             |
| DIMENSIONS, WEIGHT & POWER            |   |
| Footprint                             | 305 mm x 610 mm x 270 mm                                      |
| Weight                                | 20kg  |
| Power Consumption                     | ≤ 20 W (including detector)                                   |
| PORTABILITY                           | -<br>-  |
| Mounting                              | 305 mm x 610 mm optical breadboard; 1/4 imperial threaded     |
| Tripod                                | Optional  |
| ENVIRONMENTAL CONDITION               |   |
| Operation Temperature                 | 10°C to 40°C  |
| Storage Temparature                   | 0°C to 50°C   |
| ACCESSORIES                           |   |
| Computer                              | Not included  |
| Objective Lens                        | Included  |
| Reference Panels                      | Reflectance standard and calibration lamp                     |
|                                       |   |



## **GRAND-EOS™ HYPERSPECTRAL CAMERA**



Macro-imaging modality



Micro-imaging modality

| Spectral Range                  | 400 - 1620 nm   |  |  |
|---------------------------------|---|--|--|
|                                 | VNIR  | SWIR                                   |  |
| Spectral Resolution             | < 2.5 nm (400 - 1000 nm)  | < 4 nm (900 - 1700 nm)                 |  |
| Spatial Resolution              | < 7.5 µm  | < 12 µm                                |  |
| with 10 X microscope objective) | × 7.5 μm  |  |  |
| Camera                          | Front-illuminated interline CCD camera  | InGaAs camera deep-cooled at -80°C     |  |
| Wavelength tuning speed         | 60 ms stabilization time for 2 nm step  | 60 ms stabilization time for 5 nm step |  |
| Wavelength Absolute Accuracy    | < 0.3 nm  | < 0.6 nm                               |  |
| Visualisation Camera            | Monochrome or Color XMP camera - 2/3"   | 5.1M Progressive Color CMOS -          |  |
|                                 | 2448 x 2048 pixels  |  |  |
| Sample Holder                   | XY Manual translation stage (50 mm travel)  |  |  |
| Preprocessing                   | Image stabilization, spatial filtering, statical tools, spectrum extraction,      |  |  |
|                                 | data normalization, spectral calibration  |  |  |
| Hyperspectral Data Format       | HDF5, FITS  |  |  |
| Single Image Data Format        | HDF5, FITS, PNG, TIFF, JPG  |  |  |
| Spectrum Data Format            | HDF5, CSV, PNG, TIFF, JPG   |  |  |
| Acquisition                     | PHySpec™ control and analysis software  |  |  |
| Operating system                | Windows 7 (64 bits)   |  |  |
| Option                          | C++ SDK plugin interface included   |  |  |
| Macro-imaging modality          |   |  |  |
| Field of view                   | Optimized from 20 x 20 mm to 160 x 160 mm   |  |  |
| Lens                            | 16mm focal length VIS-NIR-SWIR objective (transmittance >90% between 400 - 1700 n |  |  |
|                                 | I   |  |  |
| Micro-imaging modality          |   |  |  |
| Microscope                      | Upright or Inverted   |  |  |
| Objectives                      | 5x, 10x (other magnifications available upon request)                             |  |  |
| Illumination                    | Broadband and monochromatic illumination available via light guide                |  |  |
| Excitation                      | 532 nm, 660 nm, 785 nm, or 808 nm lasers  |  |  |
|                                 | Other wavelengths available upon reques   | t                                      |  |

GRAND-EOS combines a hyperspectral microscopy system with a hyperspectral wide-field imaging platform, giving access to micro and macro modalities with both VNIR (400-1000 nm) and SWIR (900-1700 nm) spectral ranges. This imaging platform takes advantage Photon etc's patented filtering technology based on volume Bragg grating providing a non-polarized wavelength selection with high throughput and efficiency. This filtering method allows imaging of large field-of-view, scanning through a user defined wavelength range. Using a megapixel sensor, the acquisition of filtered images provides spectral information from million of points at the surface of the sample. The versatility of GRAND-EOS as well as its high spatial and spectral resolution makes it an ideal tool for both fundamental research or industrial applications.

## **APPLICATION EXAMPLES:**

- » Photovoltaic characterization
- >> Mineral analysis
- » Forensic
- » Food and plants sorting